

Konformitätserklärung (ISO/IEC 17050-1)
Testkörper
Supplier's declaration of conformity (in accordance with ISO/IEC 17050-1)
Test sample

Werkstoff Material	Durchmesser Diameter	Ident-Nr. Ident-no.	Dichte Density	Maß- /Formgenauigkeit Dimensional accuracy DIN 5401-2002-08	Magnetische Permeabilität Magnetic permeability
Non-Ferrous, Messing / brass CuZn36	0,80 mm	IK30049	8,4g/cm ³	Grade 100	--
		IK30077		Grade 200	--
		IK30083		Grade 200	--
		IK30090		Grade 200	--
	1,00 mm	IK30064		Grade 500	--
		IK30081		Grade 500	--
		IK30085		Grade 500	--
		IK30089		Grade 500	--
	1,20 mm	IK30060		Grade 100	--
		IK30069		Grade 100	--
		IK30084		Grade 500	--
		IK30098		Grade 500	--
	1,50 mm	IK30063		Grade 100	--
		IK30073		Grade 100	--
	1,80 mm	IK30068		Grade 500	--
		IK30070		Grade 100	--
	2,00 mm	IK30033		Grade 200	--
		IK30072		Grade 200	--
		IK30096		Grade 200	--
	2,50 mm	IK30057		Grade 200	--
		IK30075		Grade 200	--
		IK30095		Grade 200	--
	3,00 mm	IK30020		Grade 100	--
		IK30067		Grade 200	--
	3,50mm	IK30082		Grade 500	--
		IK30097		Grade 500	--



**Lebensmittelrechtliche Konformitätserklärung
nach VERORDNUNG (EU) Nr. 10/2011**

Declaration of Compliance according to REGULATION (EU) No 10/2011

Sesotec GmbH

Regener Straße 130
D-94513 Schönberg
Deutschland
Telefon: 08554 3080
Fax: 08554 2606
info@sesotec.com
www.sesotec.com

Lebensmittelbescheinigung Testkörperummantelung

Food certificate test ball cover

FDA-Konformitätserklärung / FDA-Declaration of conformity

Hiermit bestätigen wir, dass die Ummantelung unserer Testkörper aus Kunststoffen/Elastomeren, die bei bestimmungsgemäßem Gebrauch in Kontakt mit Lebensmitteln kommen können, die allgemeinen Anforderungen der FDA-Richtlinie 21 CFR §177 erfüllen.

We hereby certify that our test sample covers made of plastics/elastomers, which could come into contact with foodstuffs in the course of their intended use, meet the requirements under FDA directive 21 CFR Section 177

Enthaltene Additive entsprechen jeweils den zutreffenden FDA-Anforderungen.
Any additives that they contain meet the respectively applicable FDA requirements:

POM Blau / POM Blue (TestSticks, Testballs) - CFR 21, Part 177.2470 (Polyoxymethylene copolymer) - CFR 21, Part 178.3297 (colorants)	PE / PE (FlexStick Schlauch/FlexStick hose) - CFR 21, Part 177.1520 (Olefin polymers)	EVA Co-polymer (Flyball) - CFR 21, Part 177.1350 (EVA Co-polymers)
CARD LINE XD (Xray-Cards/Teststripes) - CFR 21, part 177.1580 (Polycarbonate)		

Die Prüfkörper mit den oben genannten Identifikationsnummern enthalten die spezifizierten Präzisionskugeln und wurden in Übereinstimmung mit den oben genannten Normen gefertigt. Wir bestätigen, dass die Lieferung mit dem Auftragseingang übereinstimmt.

The test samples with the above identification numbers contain the specified precision spheres and have been manufactured in accordance with the above standards. We confirm that the delivery corresponds to the order received.

Unbeschädigte Prüfkörper behalten bei angemessener Verwendung (Handhabung) die beschriebenen physikalischen und dimensional Eigenschaften auf unbestimmte Zeit.

Non-damaged test samples retain the described physical and dimensional properties for an unlimited period of time when used (handled) correctly.

Sesotec GmbH

29.08.2024



i. V. Dieter Sprenzinger
- Manager Quality -